



DETECTABUSE® Liquid Urine Controls for DOA Rapid Devices

- Human urine-based matrix simulates patient samples
- Recommended by the majority of DOA device manufacturers
- Stable for up to 3 years frozen, or refrigerated (unopened)
- Ready to use in 5 mL, 10mL and 20 mL vial sizes
- Available in convenient 2 pack, 6 pack or 10 pack
- Use controls to train staff in the accurate use of DOA testing devices
- Use controls to evaluate the performance of various lateral flow devices
- Formulated to work with all known drug-screening devices as an external third-party control.



Vendor Catalog#	POINT OF CARE QUALITY CONTROLS	Size
702013	KOVApoc® Pregnancy-Skreen® HCG POSITIVE (20100007) & NEGATIVE (20100005)	2X5 mL
702001	KOVAtox® DOA-SKREEN & NEGATIVE (DOA 17 drug panel including OPI300/AMP500/METH500/BE150/EDDP300/FEN50)	2X10 mL
702002	KOVAtox® DOA-SKREEN-H & NEGATIVE (DOA 17 drug panel including OPI2000/AMP500/METH500/BE150/EDDP300/FEN50)	2X10 mL
702003	KOVAtox® STAT-SKREEN 17 & NEGATIVE (DOA 17 drug panel including OPI300/AMP1000/METH1000/BE300/EDDP300/FEN50)	2X10 mL
702004	KOVAtox® STAT-SKREEN 17H & NEGATIVE (DOA 17 drug panel including OPI2000/AMP1000/METH1000/BE300/EDDP300/FEN50)	2X10 mL
702014	STAT-SKREEN® W/MDMA, OXY & BUP (19001066) & NEGATIVE (19700000) (DOA 15 drug panel, OPI300)	2X5 mL
702015	STAT-SKREEN®-H W/MDMA, OXY & BUP (19001081) & NEGATIVE (19700000) (DOA 15 drug panel, OPI2000)	2X5 mL
702018	NICOTINE (Cotinine 400ng/ml) POSITIVE (19000102) & NEGATIVE (19700000)	2X5 mL
702017	SYNTHETIC THC: (JWH-018 4 BUTANOIC ACID 100ng/mL) POSITIVE (19000959) & NEGATIVE (19700000)	2X5 mL
702016	FENTANYL: (NORFENTANYL 50 ng/mL) POSITIVE (18400933) & NEGATIVE (19700000)	2X5 mL
702028	STAT-SKREEN® W/MDMA, OXY & BUP (19002066) & NEGATIVE (19223011) (DOA 15 drug panel, OPI300)	2X20 ml
702029	STAT-SKREEN®-H W/MDMA, OXY & BUP (19002081) & NEGATIVE (19223011) (DOA 15 drug panel, OPI2000)	2X20 ml
19700000-10PK	DETECTABUSE® LIQ URINE CONTROL, NEGATIVE	10X5 mL
19000010-6PK	DETECTABUSE® LIQ URINE CONTROL, NEGATIVE	6X10 mL
20300002	ADULTER-SKREEN ,DOA Urine Validity Controls, Control A & B	2X20ml



Catalog#	STAT-SKREEN®-H	19223011, 20mL		19002081/ 19001081
	STAT-SKREEN®	19700000, 5mL		19002066/ 19001066
TEST	Constituents	Negative	Cutoff	2X
THC	Delta-9-THC-COOH	None Detected	50	100
COC	Benzoylcegonine	None Detected	300	600
PCP	Phencyclidine	None Detected	25	50
OPI/MOR	Morphine (Low Opiate)	None Detected	300	600
	Morphine (High Opiate)	None Detected	2000	4000
AMP	d-Amphetamine	None Detected	1000	2000
MET	d-Methamphetamine	None Detected	1000	2000
BARB	Secobarbital	None Detected	300	600
BZO	Oxazepam	None Detected	300	600
MTD	Methadone	None Detected	300	600
MQL	Methaqualone	None Detected	300	600
PPX	Propoxyphene	None Detected	300	600
TCA	Nortriptyline	None Detected	1000	2000
MDMA	MDMA*	None Detected	500	1000
OXY	Oxycodone	None Detected	100	200
BUP	Buprenorphine	None Detected	10	20

Constituent	CONTROL A	CONTROL B
Creatinine	Negative	>90 mg/dL (High Normal)
Nitrite	Negative	Positive
Glutaraldehyde	Positive	Negative
pH	≤4.0 (Low Abnormal)	≥9.0 (High Abnormal)
Specific Gravity	<1.005 (Low Abnormal)	>1.030 (High Abnormal)
Oxidant	Negative	Positive (PCC) Positive (H ₂ O ₂)

	Catalog #	19000010		19003020	19003030	19003066	19003081
	10ml			DOA- SKREEN	DOA- SKREEN-H	STAT- SKREEN 17	STAT- SKREEN 17H
TEST	CONSTITUENTS (ng/mL)	Negative	Cutoff	2X Cutoff	2X Cutoff	2X Cutoff	2X Cutoff
THC	Delta-9-THC-COOH	None Detected	50	100	100	100	100
COC	Benzoylcegonine	None Detected	150/300	300	300	600	600
PCP	Phencyclidine PCP	None Detected	25	50	50	50	50
OPI/MOR	Morphine	None Detected	300/2000	600	4000	600	4000
AMP	d-Amphetamine	None Detected	500/1000	1000	1000	2000	2000
MET	d-Methamphetamine	None Detected	500/1000	1000	1000	2000	2000
BARB	Secobarbital	None Detected	300	600	600	600	600
BZO	Oxazepam	None Detected	300	600	600	600	600
MTD	Methadone	None Detected	300	600	600	600	600
PPX	Propoxyphene	None Detected	300	600	600	600	600
TCA	Nortriptyline	None Detected	1000	2000	2000	2000	2000
OXY	Oxycodone	None Detected	100	200	200	200	200
MDMA	MDMA	None Detected	500	1000	1000	1000	1000
BUP	Buprenorphine	None Detected	10	20	20	20	20
BUPG	Bup. Glucuronide	None Detected	10	20	20	20	20
EDDP	EDDP	None Detected	300	600	600	600	600
FEN	Norfentanyl	None Detected	varies	50	50	50	50